
## **HCV Card Test**

For the detection of antibodies to Hepatitis C virus in human serum / plasma / whole blood



Prepare all items shown above for testing. (Timer and gloves are not provided in the test kit; they must be available)



Check expiry date on the test pouch. (If expired, do not use)



Put on gloves. (Use new gloves for each client)



Confirm product name and check

the sealing of test device pouch.

FIRST RESPONSE HCV CARD TEST

Write a client identifier on the test.




Wipe the client's finger with the alcohol swab. (Alcohol must be dry before pricking)



Remove clear protective cap of the lancet



Push lancet gently against test site. Wipe the first drop of blood and discard the lancet into sharp bin.



Gently squeeze the bulb of a specimen transfer device and touch the tip to a drop of blood. Release the bulb to draw up the blood to the specimen transfer device.



Add one drop of blood into the specimen well.



Add one drop of assay buffer into the specimen well.

Read test results at 15 Minutes. Don't interpret after 20 Minutes.

## **NEGATIVE** FIRST RESPONSE HCV = \_ c T

